



Northern Lake Service, Inc • 400 N Lake Ave • Crandon, WI 54520
800-278-1254 • www.nlslab.com

August 09, 2023

David Jadack
Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Work Order: CB07986
Received: 07/18/23

Enclosed are the results of analyses for samples received by our laboratory on 7/18/2023. If you have any questions concerning this report, please feel free to contact a client service representative at clientservices@nlslab.com.

Sincerely,

A handwritten signature in black ink, appearing to read "Ronald T. Krueger".

Ronald T. Krueger For Tracy Huber
Northern Lake Service, Inc.



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack

Reported:
8/9/23 7:25

Work Order:
CB07986

Sample Summary

Descriptions of all qualifiers listed throughout this report can be found on the Qualifiers and Definitions Page.

Lab ID	Sample	Matrix	Sample Type	Qualifiers	Date Sampled	Date Received
CB07986-01	EP500	DW			7/17/23 13:51	7/18/23 10:30
CB07986-02	EP500 Field Blank	DW			7/17/23 13:51	7/18/23 10:30

Analysis Qualifiers:

LabNumber	Analysis	Qualifier
CB07986-01	UCMR5 EPA Method 537.1	FBNA

Cancelled Tests:

LabNumber	SampleName	Analysis	Cancelled	Initials
CB07986-02	EP500 Field Blank	Perfluorinated Chemicals by EPA Method 537.1	7/21/23 13:03	CSC



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack

Reported:
8/9/23 7:25

Work Order:
CB07986

Sample Results

Sample: EP500

CB07986-01 (DW) Sampled: 07/17/23 13:51

Analyte	Result	Qualifier	MRL	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
Metals										
Lithium, Total	ND		9.0		ug/L	7/19/23 12:21	7/20/23 17:21	RAB	EPA 200.7, Rev 4.4	1
<i>Surrogate: Yttrium 200.7 ISTD</i>										
	95%		Limits: 60-125%			7/19/23 12:21	7/20/23 17:21	RAB	EPA 200.7, Rev 4.4	
Semi-Volatiles										
perfluorobutanoic acid (PFBA)	ND		0.0050		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoro-3-methoxypropanoic acid (PFMPA)	ND		0.0040		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoropentanoic acid (PFPeA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorobutanesulfonic acid (PFBS)	0.0030		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoro-4-methoxybutanoic acid (PFMBA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	ND		0.020		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorohexane sulfonic acid (4:2FTS)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorohexanoic acid (PFHxA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoropentanesulfonic acid (PFPeS)	ND		0.0040		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.0050		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoroheptanoic acid (PFHpA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorohexanesulfonic acid (PFHxS)	0.0091		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorooctane sulfonic acid (6:2FTS)	ND		0.0050		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoroheptanesulfonic acid (PFHpS)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorooctanoic acid (PFOA)	ND		0.0040		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorononanoic acid (PFNA)	ND		0.0040		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorooctanesulfonic acid (PFOS)	ND		0.0040		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	ND		0.0020		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorodecanoic acid (PFDA)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorodecane sulfonic acid (8:2FTS)	ND		0.0050		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluoroundecanoic acid (PFUnA)	ND		0.0020		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack

Reported:
8/9/23 7:25

Work Order:
CB07986

Sample Results (Continued)

Sample: EP500 (Continued)

CB07986-01 (DW) Sampled: 07/17/23 13:51

Analyte	Result	Qualifier	MRL	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
Semi-Volatiles (Continued)										
11-chloroeicosafuoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		0.0050		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
perfluorododecanoic acid (PFDa)	ND		0.0030		ug/L	7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	1
Surrogate: (EIS) MPFBA	92%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M5PFPeA	96%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M3PFBS	93%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M2-4:2FTS	136%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M5PFHxA	90%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M4PFHpA	89%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M3PFHxS	96%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M2-6:2FTS	111%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M8PFOA	88%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M9PFNA	88%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M8PFOS	97%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M2-8:2FTS	106%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M6PFDA	86%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M7PFUDa	90%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) MPFDoA	91%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
Surrogate: (EIS) M3HFPODA	84%		Limits: 50-200%			7/31/23 5:16	7/31/23 17:06	RAW	EPA 533	
n-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		0.0060		ug/L	7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	1
N-ethyl perfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		0.0050		ug/L	7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	1
perfluorotridecanoic acid (PFTTrDA)	ND		0.0070		ug/L	7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	1
perfluorotetradecanoic acid (PFTa)	ND		0.0080		ug/L	7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	1
Surrogate: (SURRE) C13-PFHxA	93%		Limits: 70-130%			7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	
Surrogate: (SURRE) C13-HFPODA	93%		Limits: 70-130%			7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	
Surrogate: (SURRE) C13-PFDA	92%		Limits: 70-130%			7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	
Surrogate: (SURRE) d5-NEtFOSAA	91%		Limits: 70-130%			7/19/23 10:00	7/21/23 20:13	RAW	EPA 537.1, Rev 2.0	

Sample: EP500 Field Blank

CB07986-02 (DW) Sampled: 07/17/23 13:51

Analyte	Result	Qualifier	MRL	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
---------	--------	-----------	-----	-----	-------	---------------	---------------	---------	--------	---------------



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack

Reported:
8/9/23 7:25

Work Order:
CB07986

Sample Results (Continued)

Sample: EP500 Field Blank (Continued)

CB07986-02 (DW) Sampled: 07/17/23 13:51

Analyte	Result	Qualifier	MRL	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
Semi-Volatiles										
perfluorobutanoic acid (PFBA)	ND		0.0050		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoro-3-methoxypropanoic acid (PFMPA)	ND		0.0040		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoropentanoic acid (PFPeA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorobutanesulfonic acid (PFBS)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoro-4-methoxybutanoic acid (PFMBA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
nonafluoro-3,6-dioxaheptanoic acid (NFDHA)	ND		0.020		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorohexane sulfonic acid (4:2FTS)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorohexanoic acid (PFHxA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoropentanesulfonic acid (PFPeS)	ND		0.0040		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.0050		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoroheptanoic acid (PFHpA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorohexanesulfonic acid (PFHxS)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorooctane sulfonic acid (6:2FTS)	ND		0.0050		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoroheptanesulfonic acid (PFHpS)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorooctanoic acid (PFOA)	ND		0.0040		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorononanoic acid (PFNA)	ND		0.0040		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorooctanesulfonic acid (PFOS)	ND		0.0040		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	ND		0.0020		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorodecanoic acid (PFDA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
1H,1H, 2H, 2H-perfluorodecane sulfonic acid (8:2FTS)	ND		0.0050		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluoroundecanoic acid (PFUnA)	ND		0.0020		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
11-chloroicosafafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		0.0050		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
perfluorododecanoic acid (PFDoA)	ND		0.0030		ug/L	8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	1
Surrogate: (EIS) MPFBA	99%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) MSPFPeA	100%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	



Eau Claire Water/Wastewater Utilities
 1000 Ferry Street
 Eau Claire, WI 54703

Project: UCMR5 Testing SE2
 Project Number: UCMR5 Testing
 Project Manager: David Jadack

Reported:
 8/9/23 7:25

Work Order:
 CB07986

Sample Results (Continued)

Sample: EP500 Field Blank (Continued)

CB07986-02 (DW) Sampled: 07/17/23 13:51

Analyte	Result	Qualifier	MRL	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
Semi-Volatiles (Continued)										
Surrogate: (EIS) M3PFBS	97%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M2-4:2FTS	105%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M5PFHxA	96%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M4PFHpA	98%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M3PFHxS	99%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M2-6:2FTS	98%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M8PFOA	98%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M9PFNA	100%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M8PFOS	100%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M2-8:2FTS	105%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M6PFDA	101%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M7PFUdA	95%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) MPFDaA	92%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	
Surrogate: (EIS) M3HFPODA	89%		Limits: 50-200%			8/2/23 5:23	8/2/23 15:48	RAW	EPA 533	



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack

Reported:
8/9/23 7:25

Work Order:
CB07986

List of Certifications

Code	Description	Number	Expires
1	EPA Laboratory ID No.	WI00034	1/1/26



Eau Claire Water/Wastewater Utilities
1000 Ferry Street
Eau Claire, WI 54703

Project: UCMR5 Testing SE2
Project Number: UCMR5 Testing
Project Manager: David Jadack


Reported:
8/9/23 7:25

Work Order:
CB07986

Qualifiers and Definitions

Item	Definition
FBNA	The field sample had no detects, therefore the corresponding trip blank/field reagent blank was not analyzed.
ND	Analyte NOT DETECTED at or above the LOD or MRL.
LOD	Limit of Detection.
LOQ	Limit of Quantitation.
NA	Not Applicable.
Dry	Dry Weight Basis.
Wet	Wet Weight Basis.
% Dry	Equal to: (mg/kg dry) / 10000.
1000 ug/L	Equal to: 1 mg/L.
MCL	Maximum Contaminant Levels for Drinking Water Samples. Shaded results indicate >MCL.
RPD	Relative Percent Difference.
%REC	Percent Recovery.
Source	Sample that was matrix spiked or duplicated.

All LOD/LOQs adjusted to reflect preparation volumes, dilutions, and/or solids content.



Northern Lake Service Inc.
 ENVIRONMENTAL ANALYTICAL LABORATORY
 400 N Lake Ave | Crandon, WI 54520 | 800-278-1254 | www.nlslab.com

UCMR5 Sample Collection and Chain of Custody Record
 Eau Claire Water/Wastewater Utilities -- UCMR5 Testing SE2

Bottle Order Number: C000124

(Client please fill in date/time)

Sample Point ID Code	Collection Date/Time	200.7	533	537.1
EP500	7-17-2023 1:51 PM	X	X	X
EP500 Field Blank	7-17-2023 1:51 PM		X	X

The EPA Unregulated Contaminant Monitoring Program has specific sample receipt temperature requirements. Samples received within 48 hours from collection must be received at < or = 10 C. Samples received greater than 48 hours from collection must be received at < or = 6 C

(Client please sign)

Collected by (signatures) *[Signature]*
 Method of Transport **Fed Ex**
 Sample Collection Comments (Optional)

(To be filled out by lab upon arrival)

	200.7	533	537.1
Received at NLS by (Signature) <i>[Signature]</i>	Date/Time 7/18/23 10:30	Rec Temp 1.8	
Remarks and other Information FedEx			

IMPORTANT: To meet regulatory requirements, this form must be completed in detail and included in the cooler containing the samples described

Q

For Lab Use Only										
Method 533	S1B1	S1B2	S2B1	S2B2	S3B1	S3B2	S4B1	S4B2	S5B1	S5B2
pH (6-8)	7.38	7.48	7.21	7.37						
Res Chlorine (<0.1)	CL	CL	CL	CL						
Method 537.1	S1B1	S1B2	S2B1	S2B2	S3B1	S3B2	S4B1	S4B2	S5B1	S5B2
pH (6-8)	7.47	7.47	7.48	7.49						
Res Chlorine (<0.1)	CL	CL	CL	CL						
Method 200.7	S1	S2	S3	S4	S5					
Acidified	/									

CB07986

